

Surface water chemical Quality Tests (SQT)

Scope:



Physical & Sensory Parameters:

Colour (Appearance & True Colour Units), Odour, pH, Turbidity; Total suspended solids (TSS), Electrical Conductivity (EC) & Total Dissolved Solids (TDS) by gravimetry (TDSG).



Chemical Parameters:

Alkalinity, Total hardness of water (THW), Calcium (Ca) & Magnesium (Mg); Ammoniacal-Nitrogen ($\text{NH}_3/4\text{-N}$), Nitrite (NO_2), Nitrate (NO_3); Chloride, Sulfate, Phosphate, Sodium, Potassium, Total Oil & Grease (TOG); Sodium Adsorption Ratio (SAR).

Total parameters: $7 + 12 = 19 + 1$ computed (Sodium Adsorption Ratio)

Rationale:

Surface water is any body of water above ground, including streams, rivers, reservoirs, lakes, tanks, and ponds¹. Most water utilities source raw water from surface water bodies. For example; the Hyderabad metro water supply is sourced from reservoirs built across several rivers, namely; Musi, Manjeera, Krishna and Godavari. After satisfying drinking and domestic water supply needs, surface water is used for several important purposes, such as irrigation, recreation, fishery, etc. Pristine water stored in a protected reservoir, can be used for drinking with minimal treatment, would be a good aquatic habitat, useful for fishery and suitable for irrigation. In addition, water reservoirs provide excellent opportunities for recreation, by way of a scenic water front, healthy for bathing and swimming, or for that matter sailing, boating etc.

Dynamic interaction of atmosphere, rainfall pattern, and geochemical conditions of drainage basins are natural determinants of water quality characteristics of freshwater bodies. If surface waters were totally unaffected by human activities, up to 90-99 per cent of global freshwaters, would have natural chemical concentrations suitable for aquatic life and most human uses (Chapman, 1996). Continuous flow in streams and rivers and large area of surface water bodies do provide some natural regenerative capacity against pollution due to exposure to animal and human activities. However, various factors such as dynamic interaction of changes in natural flow, manmade obstructions, inadequate protection, encroachments, urban run-off, inadequately treated industrial effluents and many more factors can affect surface water bodies, which in turn may affect suitability for various intended use. The Central Pollution Control Board uses the concept of designated-best-use for monitoring and assessment of water quality in surface water bodies. A particular water body may have several uses and the quality criteria for each would vary. For any given water body meant for several purposes, the use which demands highest quality of water is called its

¹ Although, oceans, seas, salt-water lakes and lagoons are surface water bodies, we restrict here to fresh-water bodies only.

designated-best-use. Five quality classes are identified adopting a minimal set of primary water quality criteria. (Table 1).

Table 1: Use based classification of surface waters in India.

Quality Class	Designated Best Use	Primary Water Quality Criteria			
		pH	Dissolved Oxygen	Biochemical O ₂ Demand	MPN (Total coliforms)
A	Drinking water source without any treatment, except for chlorination.	6.5-8.5	≥ 6 mg/L	≤ 2mg/L	≤ 50 cfu/100ml
B	Organised outdoor bathing	6.5-8.5	≥ 5 mg/L	≤ 3mg/L	≤ 500 cfu/100ml
C	Drinking water source with conventional treatment	6.0-9.0	≥ 4 mg/L	≤ 3mg/L	≤ 5000 cfu/100ml
D	Propagation of wildlife and fisheries	6.5-8.5	≥ 4 mg/L	Ammonia ≤ 1.2 mg/L	
E	Irrigation, industrial cooling, and controlled disposal.	6.0-8.5	EC ≤ 2250 μS/cm; Sodium Adsorption Ratio < 26; Boron < 2 mg/L		

Source: Based on Table 1 in CPCB. 2008. Guidelines for Water Quality Management. Government of India, Ministry of Environment & Forests - Central Pollution Control Board (CPCB), New Delhi.

SQT is SBP minus the bacteriological parameters. In SQT we have packaged a broader set of physical and chemical parameters, for routine monitoring of surface water quality to determine prima-facie suitability for various common purposes and sources of pollution, if any. The natural acid-base balance of a water body can be affected by industrial effluents and atmospheric deposition of acid-forming substances. Changes in pH can indicate the presence of certain effluents, particularly when continuously measured and recorded, together with the electrical conductivity. Excess of TSS, and turbidity affects fishery, suitability as a source of drinking water, bathing and recreational use, and may be due to impact of untreated sewage, urban run-off, agricultural activities in the foreshore, and certain industrial effluents. EC and TDS indicate salinity, which is an important consideration for

Surface water chemical Quality Tests (SQT)



irrigation & hydroelectric power plants. Hazardous chemical waste, food processing, textile, chemical and pharmaceutical effluent, can substantially increase EC and TDS.

In the past, people have ordered SQT to check if an urban water body is polluted, whether a tank or pond is suitable for fishery, and if a source is good irrigation water for agriculture / horticulture. In situations where

contamination of a surface water body is obvious, people have brought samples for to assess physical and general chemical parameters as indicators of the level of pollution.

Sample - Collection, Storage & Transportation:

Follow methods of sampling specified in IS 3025 part 1 : 1987 for chemical tests. Sample should be representative of the water to be tested and should be collected with utmost care to ensure that no contamination occurs at the time of collection or prior to examination by the laboratory. Choice of container and appropriate preparation is important.

Step-1: Gather all that you need for collection of water sample:

Two litre sample collected in a clear or amber colour clean polypropylene can or bottle is required for physical and chemical tests. You need two 1 Litre clean dry polypropylene bottles (2 x CBWS/ABWS), two black or dark colour polythene bags (small garbage bag will do) to minimise exposure of samples to sunlight, ice packs to keep the sample bottles cool during transport and a carry bag for convenient transport.

If it is not feasible for you to collect the specified sample collection bottles from the laboratory and you must collect samples, freshly emptied packaged drinking water bottles or new PET water bottles and cans may be used, in that order. Do not use empty beverage bottles such as ThumsUp, Maaza, Sprite, Coca Cola. Nutrients and other residues in such bottles may affect some physical and chemical parameters.

Step-2: Identify sampling point and time:


Quality of water often varies in different parts of surface water bodies. Therefore, a single sampling-point is usually not enough. The number of sampling-locations and -depths depend on size of the water body and study objectives. However, a single sample coupled with contextual data can help determine prima-facie suitability for intended use and yield useful clues about possible sources of pollution, if any. If you plan to take a single sample, prefer the deeper ends, or clear areas where water is more than one meter. If feasible, approach the sampling point by a boat or from an accessible platform on water. Otherwise, slowly wade in towards the spot, up to about your waist deep, taking care to minimise disturbance of the bottom, so that you do not kick up any sediment to rise to surface. After reaching the sampling spot, wait for a few minutes for kicked up sediments around you to settle down. Then extend your arm to collect sample from a spot minimally affected by your entry. If this is not feasible, identify a spot on the bank from where you can draw water using a pole and a clean bucket.

For nalas, streams and rivers, identify a spot upstream of any bridge, culvert, crossing. Collect sample slightly below surface of flowing water, while avoiding bottom of the stream. If depth of stream permits, immerse the bottle completely about 4 inches deep. If stream is too shallow to immerse the bottle fully, collect as much as possible, being very careful not to touch the bottom where sediments can be disturbed and make sure no surface film flows into the bottle.


Step-3: Collect sample.

- a. Always have an assistant to standby, just in case you need help as you wade into the waterbody or stream, to help draw water using etc. Both you and the assistant should wash both hands with soap and water, either at a distance from

the intended sampling point, or by drawing a bucket of water to the bank for this purpose.



- b. Ready a clean and dry bucket and suitable pole and/or rope if direct collection by immersion is not feasible.
- c. Label the sample collection bottles and place it within easy reach, but do not open at this stage. Have ice packs ready.
- d. Remember that mouth of the dipped bottle should not go too close to the lake bottom. The ideal depth zone for intake of grab sample would be at least 20 cm above lake floor and 20 cm below surface. Thus, the depth of water at the sampling point should be at least 2ft, so that there is about 10 cm of intake zone.
- e. Carefully wade into the lake and stop at a point where water is knee deep. Stand facing the lake. Stand still for some time, until kicked up sediments settle down.
- f. Face towards the lake centre (deeper area). You will need to bend forward to collect sample from a point that has not been disturbed by your entry into the lake.
- g. Immerse the open bottle by hand into surface water and allow water to slowly run into the bottle minimizing turbulence. If this is not feasible, use clean bucket and pole/rope to draw water and pour into sample collection bottles, close cap tightly.
- h. Hold the bottle towards deeper part of the lake, tilt it downwards and plunge the bottle opening under the water surface, until your elbow touches water surface. Turn the submerged bottle gradually as your elbow makes 45° angle with respect to water surface, and the opening pointed away from you, towards deeper areas of the lake. Fill the bottle slowly, allowing air inside the bottle to escape gradually. As the bottle fills more and more, gradually turn the bottle up until its mouth looks up vertically upwards, while still under water.
- i. Put sample bottles in dark bags, wrap ice packs and place in carry bag.



Step-4: Transport to laboratory: Preferably in 6 hours.

Step-5 Store sample, if required:

If immediate transport is not feasible, store the sample inside the regular chamber (not the freezer compartment) of a refrigerator until you are ready to transport it to the Laboratory, and definitely within 24 hours from the time of collection.

Information About Source, Context, Intended Use & Concerns:

Provide as much detail as you can about the water body, sampling point, activities & environment around the sampling point. Click some photos and record GPS coordinates, if you can. Mention about intended use of the water, the reason why you are ordering the test, as well as doubts and concerns, if any. Such information helps interpretation of test results. Occasionally, the IHS Laboratory may contact you for clarifications and additional information about the source and its environment, to help interpretation of test results.

Test Method & Duration:

Organoleptic, physical and chemical characteristics of water sample are tested according appropriate parts of the IS3025 and/or American Public Health Association (APHA). Report will be available in 2 to 3 days.

To pick up sample collection bottle and/or schedule collection of samples:
Email: ihslab@ihs.org.in; WhatsApp: +919848011251; Call:040-23211013/4



For various water quality test packages: <https://www.ihs.org.in/lab/wqt.html>
& To download complete water quality test catalogue in tabular form, click:
<https://www.ihs.org.in/lab/wqt/pdf/IHSLabWaterQualityTestsCatalogue.pdf>

References:

- CPCB. 2008. Guidelines for Water Quality Management. Government of India, Ministry of Environment & Forests - Central Pollution Control Board (CPCB), New Delhi.
- Chapman Deborah. 1996. Water Quality Assessments - A Guide to Use of Biota, Sediments and Water in Environmental Monitoring. London: published on behalf of WHO by F & FN Spon.
- IS 3025 Part 1: 1987. Methods of Sampling and Test (Physical and Chemical) for Water and Wastewater. Part 1 Sampling. New Delhi: Bureau of Indian Standard (BIS); Indian Standard. Reaffirmed 2003.

----- © 2022-2023, The Institute of Health Systems, Hyderabad, India. -----